


Recommended Resupply Schedule

Replace your CPAP supplies as recommended to maintain the most effective sleep apnea treatment.



2 weeks

Disposable Filters

Replace disposable filters to prevent fine particles from entering your airway.



2 weeks

Mask Cushions and Pillows


Replace mask cushions and pillows to help stop mask leaks.



3 months

Face Masks

Replace your face mask to prevent discomfort and inefficient CPAP treatment.



3 months

Tubing


Replace tubing to ensure safe and clean air quality.



6 months

Headgear and Chin Straps

Get new headgear to replace any stretched and worn out products, which can cause mask leaks.



6 months

Non-Disposable Filters


Protect yourself from particle buildup and maintain good air quality with non-disposable filters.



6 months

Humidifier Water Chamber

Prevent mineral buildup from getting in your airway with a new humidifier chamber.



5 years

CPAP/APAP Machine

Replace your CPAP machine to stay up-to-date on the most effective therapy options.